



Český metrologický institut
Okružní 31, 638 00 Brno



Notified Body
No. 1383

tel. +420 545 555 111, fax +420 545 222 728, www.cmi.cz

EC-TYPE EXAMINATION CERTIFICATE

Number: TCM 142/14 - 5171

Page 1 from 9 pages

In accordance: with Directive 2004/22/EC of the European Parliament and of the Council as amended implemented in Czech Republic by Government Order No. 464/2005 Coll. as amended that lays down technical requirements on measuring instruments.

Manufacturer: ŞENSOYLAR Klape Tesizat Malz. San. ve Tic. Ltd. Şti.
Bakırcılar ve Prinççiler Sanayi Sitesi Menekşe Caddesi No:13
İstanbul / Beylikdüzü
Turkey

For: water meter - multi jet, dry dial
type: ERCIYES MJD-xx-T
Accuracy class: 2
Temperature class: T30 or T50

Valid until: 22 June 2024

Document No: 0511-CS-A049-14

Description: Essential characteristics, approved conditions and special conditions, if any, are described in this certificate.

Date of issue: 23 June 2014

Certificate approved by:




RNDr. Pavel Klenovský

This certificate was issued according to module B - type examination according to annex B to Directive 2004/22/EC of the European Parliament and of the Council or point 3 of annex 2 to Government Order No. 464/2005 Coll., respectively.